



## Conductivity meter AX46010001

URL:<https://www.sxplc.com/conductivity-meter-ax46010001>

### Product data sheet

Control Function – AX460 Only

Controller Type

P, PI, PID (Configurable)

Control Outputs

Analog

Current output control (0 to 100%)

Time proportioning cycle time

5 to 60s, programmable in 1s steps displayed with doP or  
roP

Pulse frequency

1 to 120 pulses per minute

Controller action

Reverse, direct or bi-directional (programmable)

Proportional band

0.1 to 1000%, programmable in 0.1% increments

Integral action time (Reset)

1 to 7200s in 1s steps, 0 = Off

Derivative

0.1 to 1000s in 0.1s increments

Auto/Manual

User-programmable

---

